MEBEP TECH(HK) Co., Limited
Email: sales@mebep.com Website: www.mebep.com
Tel: +86-755-86134126 WhatsApp/Facebook/Twitter: +86-189-22896756

Quantitative Diagnostic Kit for Human Immunodeficiency Virus type 1 (HIV-1) RNA

Product Number: DTL0416

Shipping and Storage

1. Store below 30℃. It is valid for 12 months.
2. Transport at normal temperature, not suggested over 14 days.
3. Opened but not completely used the all components should be stored at (-20±5)℃. It is recommended to separate in PCR tubes before refrigeration to avoid repeated freezing and thawing of all reagents next time. It is not recommended to repeat the freeze-thaw cycle more than 7 times.
4. Date of manufacture and term of validity: see the label.

Component

Component	48T
HIV-1 RT-PCR Master Mix	Lyophilized powder ×1 Bottle
HIV-1 quantitative standard substance 1	Lyophilized powder ×1 Tube
HIV-1 quantitative standard substance 2	Lyophilized powder ×1 Tube
HIV-1 quantitative standard substance 3	Lyophilized powder ×1 Tube
HIV-1 quantitative standard substance 4	Lyophilized powder ×1 Tube
Positive Control	100μL
Negative Control	1mL
Redissolved Diluent	1.5mL

Note: Do not mix reagents from different batches.

The reaction system is lyophilized powder that contains all components required for fluorescence PCR, including Taq enzyme, primers, probes,reverse transcriptase, dNTPs, and Mg^{2+}.

Description

This kit uses multiple fluorescent PCR detection technology to achieve quantitative detection of HIV-1 RNA in the sample to be tested by using a pair of specific primers and a specific fluorescent probe for HIV. The kit is provided with an Internal Control, which can monitor whether there is PCR inhibitor in the sample to be tested by detecting whether the internal control is normal or not, so as to avoid false negative PCR.

Application

This kit is suitable for in vitro quantitative detection of Human Immunodeficiency Virus type 1 （HIV-1） nucleic acid in serum samples.The test results can be used for auxiliary diagnosis of HIV-1 infection and evaluate the efficacy of antiviral therapy and monitor recurrence. The test results are for clinical reference only and should not be used as the sole criterion for clinical diagnosis. This kit should not be used for blood screening.

AIDS is a chronic disease caused by human immunodeficiency virus (HIV), which destroys the immune capacity of the human body and causes the immune system to lose its resistance. It has the characteristics of strong infectivity, wide prevalence and higher incidence. Common clinical symptoms include persistent extensive lymphadenopathy, fever, night sweats, severe fatigue, weight loss, etc.

Applicable instruments

Real-time fluorescence PCR instrument with FAM,VIC channels.

Specimen collection

1. Applicable sample type: Serum specimens.
2. Sample collection: Use a sterile syringe to draw 2mL of the subject's venous blood and inject it into a sterile centrifuge tube. Leave it at room temperature for no more than 4 hours. Centrifuge at 2000rpm for 5 minutes. Take the upper serum (be careful not to bring in red blood cells) and transfer it to another sterile centrifuge tube for use.
3. Sample storage and transportation: The collection or processing sample should not exceed 24 hours under the conditions of 2℃ ~ 8℃. If long-term preservation is needed, it should be stored below -70℃, and the freezing fusion should not exceed 3 times.

Protocol

1. Reagent preparation:

Take out the HIV-1 RT-PCR Master Mix, open each bottle cap according to the arrow direction of the aluminum-plastic cover, add 960μL of Redissolved Diluent, strongly mixed on the vortex for more than 1 minute, then stand for 30 ~ 60 seconds until the liquid is clear and transparent. Subpackage it into PCR reaction tubes according to 20μL/ tube.

2. Nucleic acid extraction:

This kit is not included for Nucleic Acid(NA) extraction reagent.

Commercially available extraction kits that have been shown to generate highly purified RNA when following manufacturer's recommended procedures for sample extraction are applicable.

If the extracted RNA is not used immediately, it should be stored below -20℃. For long-term storage, it should be stored below -80℃ and avoid repeated freezing and thawing.

Note: The Negative Control and the Positive Control does not require nucleic acid extraction.

The HIV-1 quantitative standard substances 1~4 needs to be redissolved with 100μL of Redissolved Diluent and mixed well before use separately.

3. Add sample:

The correspond substances were added to that above PCR reaction tubes according to the following table:

Type	Add sample description
Testing Sample	Add 5μL of the extract prepared in step 2 to the reaction tube, and close the tube cover.
Negative Control/ Positive Control	Add 5μL of negative control and positive control to the reaction tube, and cover the tube tightly.
HIV quantitative standard substances 1,2,3,4	Add 5μL of HIV-1 quantitative standard substances 1,2,3,4 respectively, to the PCR reaction tube, and cover the tube tightly.

The total reaction volume is 25μL.

After adding the sample, the PCR reaction tubes should be centrifuged for 15s on a palm centrifuge and then delivery to the nucleic acid amplification region. If bubbles are found, the tube wall should be gently flicked to remove bubbles and centrifuged again.

4. PCR amplification:

Place the reaction tube in the automatic fluorescent PCR instrument, set the Negative control, Positive control, and test sample parameters to perform PCR experiment according to the operating instructions of the instrument, and record the corresponding sample name.

Select FAM channel to detect HIV-1 RNA, select VIC channel to detect the IC. Set the Reaction Volume per Well to 25μL.

(Note: For ABI series instruments, select 'None' under 'Quencher', and select 'None' as the dye to use as the passive reference.)

Recommended reaction program setting:

Step	Cycles	Temperature	Time	Collect fluorescence signal
1	1 cycle	50℃	10min	No
2	1 cycle	95℃	2min	No
3	45 cycles	95℃	15sec	No

		60℃	30sec	Yes

5. Result analysis:

After the reaction is completed, the results are automatically saved.

The Start value, End value and Threshold value of the Baseline should be adjusted according to the analyzed image (the user can adjust it according to the actual situation, the Start value can be set at 3~15, the End value can be set at 5~20, the amplification curve of the negative control should be adjusted to be flat or below the threshold line).

Click Analyze for analysis, make the parameters meet the requirements in the following '6.Quality control', and then go to the Plate window to record the Ct value.

6. Quality control

Negative control: FAM and VIC detection channels have no amplification curves,in the Reports interface, the Ct column displays 'Undet'.

Positive control:The FAM and VIC channels have obvious S-shaped amplification curves, and the Ct value is ≤32.00. HIV quantitative standard substances 1~4:FAM channel has obvious amplification curve, and the standard curve correlation coefficient $|r| \geq 0.98$.

The above requirements must be met at the same time in the same experiment. Otherwise this experiment is invalid and needs to be repeated.

Explanation of Test Result

1. If the Ct value of the VIC channel of the test sample is >35.00, the test result of the sample is invalid, and the cause should be found and eliminated. Samples need to be recollected and the experiment repeated.
2. For samples with measured value $<5.00\times10^2$ copies/mL or Ct value was 'Undet', and the Internal Control detection is positive and Ct value≤32.00, it was reported as lower than the detection limit of the kit. If the Internal Control is abnormal, the test result of this sample is invalid, and the test is repeated for this sample.
3. For samples with measured values between 5.00×10^2copies/mL~1.00×10^8 copies/mL, report the corresponding determination results.
4. For samples with measured value $>1.00\times10^8$ copies/mL, the Internal Control detection is positive and Ct value≤32.00, It is recommended to re-measure after diluting with 1×TE, calibrate the results according to the dilution factor, report the calibrated value, and indicate "reportable concentration",and the measurement results are for reference only.

Limitation

1. Sample detection results are related to sample collection, processing, transportation and preservation quality.
2. There is no control of cross -pollution during the sample extraction process, and false positive results will occur.
3. Positive control and leakage of amplification products can lead to false positive results.
4. The genetic mutations and reorganizations during epidemics can lead to false negative results.
5. Different extraction methods have differences in extraction efficiency, which will lead to false negative results.
6. Reagent transportation, improper preservation, or inaccurate reagent preparation reagent detection performance decreases, and the results of false negative or quantitative detection occur.
7. The results of this test are for reference only. If the diagnosis must be confirmed, please combine clinical symptoms and other test methods.

Performance Parameters

1. Minimum detection limit: The minimum detection limit of this reagent is 5.00×10^2 copies/mL.
2. Linear range: 5.00×10^2 copies/mL~1.00×10^8 copies/mL.
3. Precision: Repeat detection of the enterprise precision reference product 10 times, and the coefficient of variation (CV, %) value of detected concentration logarithm is ≤5.00%.
4. Compliance rate of negative/positive reference products: The compliance rate of negative reference products in enterprise

reference is 100%, and the compliance rate of positive reference products is 100%.

Note

1. Please read the instructions of this kit carefully before the experiment, and strictly follow the operation steps.
2. Before the test, please be familiar with and master the operation method and precautions of various instruments to be used, and carry out quality control for each experiment.
3. The reaction solution should be stored away from light.
4. Try to avoid bubbles in the reaction, and the tube cover needs to be tight.
5. Use disposable heads, disposable gloves and special work clothes in each district.
6. Sample processing, reagent preparation, and samples need to be performed in different areas to avoid cross -pollution.
7. After the experiment is completed, use 10% hypochlorite or 75% alcohol or ultraviolet light to treat the workbench and pipette.
8. All items in the kit should be treated as pollutants and processed in accordance with the "Biological Safety General of Microbiological Biomedical Laboratory".