

Bovine Theileria annulata ELISA Kit(Qualitative)

Product Number: ELK066

Shipping and Storage

1. Reagent kit storage: 2-8°C.
2. Validity period: 6 months.

Component

Component	96T
30× concentrated wash buffer	20mL × 1
Enzyme conjugate	6mL × 1
Microplate (Coated)	12 wells × 8 strips
Sample diluent	6mL × 1
Chromogen substrate A	6mL × 1
Chromogen substrate B	6mL × 1
Stop solution	6mL × 1
Positive control	0.5mL × 1
Negative control	0.5mL × 1

Description

This test kit uses the double antigen sandwich method to determine the expression of bovine circular Taylor disease antibody (T. annulata Ab) in specimens. Using purified antigen coated microplates to prepare solid-phase antigens, they can bind to T. annulata Ab in the sample. After washing to remove unbound antibodies and other components, they can then bind to HRP labeled antigens to form antigen antibody antigen complexes. After thorough washing, TMB substrate is added for color development. TMB is converted to blue under the catalysis of HRP enzyme and to the final yellow under the action of acid. Measure the absorbance (OD value) at a wavelength of 450nm using an enzyme-linked immunosorbent assay (ELISA) reader, and compare it with the CUTOFF value to determine the presence or absence of T. annulata Ab antibodies in the specimen.

Application

This kit is designed for the detection of antibodies against Theileria annulata (T. annulata Ab) in bovine serum, plasma, and related liquid samples.

Specimen requirements

1. Extract the specimen as soon as possible after collection, according to relevant literature, and conduct experiments as soon as possible after extraction. If the experiment cannot be conducted immediately, the specimen can be stored at -20°C, but repeated freezing and thawing should be avoided.
2. Samples containing NaN₃ cannot be detected because NaN₃ inhibits the activity of horseradish peroxidase (HRP).

Protocol

1. Number: Number the corresponding micropores of the sample in sequence. Each plate should have 2 negative control wells, 2 positive control wells, and 1 blank control well (blank control wells do not contain samples or enzyme-linked immunosorbent assay reagents, and all other steps are the same)
2. Sample addition: Add 50µL of negative control and positive control to the negative and positive control wells respectively. Then add 40µL of sample diluent to the test sample well, followed by 10µL of the test sample. Add the sample to the bottom of the enzyme-linked immunosorbent assay plate well, avoiding touching the well wall as much as possible. Gently shake and mix

For Research Use Only

well,

3. Incubation: Cover the plate with a sealing film and incubate at 37°C for 30 minutes.
4. Solution preparation: Dilute 30 times the concentrated washing solution with distilled water and set aside for later use
5. Washing: Carefully remove the sealing film, discard the liquid, shake dry, fill each hole with detergent, let it stand for 30 seconds, then discard. Repeat this process 5 times and pat dry.
6. Enzyme addition: Add 50µL of enzyme labeled reagent to each well, except for blank wells
7. Incubation: Follow the same procedure as in 3.
8. Washing: The operation is the same as 5.
9. Color development: Add 50µL of color developing agent A50 to each well, then add 50µL of color developing agent B50, gently shake and mix, and develop color at 37°C in the dark for 15 minutes
10. Termination: Add 50µL of termination solution to each well to terminate the reaction (at this point, the blue color turns yellow).
11. Measurement: Zero the blank well and measure the absorbance (OD value) of each well in sequence at a wavelength of 450nm. The measurement should be conducted within 15 minutes after adding the termination solution.

Calculation and result judgment

Experimental effectiveness: The average value of positive control wells is ≥ 1.00 ; Negative control mean ≤ 0.20 ;

CUT OFF calculation: $CUT = \text{mean value of negative control wells} + 0.15$;

Negative determination: Samples with OD values less than the critical value (CUT OFF) are negative for T. annuata Ab antibodies;

Positive determination: Samples with OD values \geq the critical value (CUT OFF) are considered positive for T. annuata Ab antibodies.

Note

1. Strictly follow the instructions for operation, and components from different batches of this reagent must not be mixed.
2. The kit should be balanced at room temperature for 15-30 minutes before being taken out of the cold storage environment. If the enzyme coated plate is not used up after opening, the Flat noodles should be stored in a sealed bag.
3. Concentrated washing solution may precipitate crystals. When diluting, it can be dissolved by heating in a water bath, and washing does not affect the results.
4. The sealing film is only for one-time use to avoid cross contamination.
5. Please store the substrate away from light.
6. The determination of experimental results must be based on the reading of the enzyme-linked immunosorbent assay (ELISA) reader. When using dual wavelength detection, the reference wavelength is 630nm
7. All samples, detergents, and various waste materials should be treated as infectious agents. The termination solution is 2M sulfuric acid, and safety must be taken into account when using it.